			ASK	
	Application No.	Applicant(s)	Applicant(s)	
Notice of Allowability	10/625,437	IATROU ET AL.	IATROU ET AL.	
	Examiner	Art Unit		
	Chih-Cheng Glen Kao	2882		
The MAILING DATE of this communication appears on the cover sheet with the correspondence address All claims being allowable, PROSECUTION ON THE MERITS IS (OR REMAINS) CLOSED in this application. If not included herewith (or previously mailed), a Notice of Allowance (PTOL-85) or other appropriate communication will be mailed in due course. THIS NOTICE OF ALLOWABILITY IS NOT A GRANT OF PATENT RIGHTS. This application is subject to withdrawal from issue at the initiative of the Office or upon petition by the applicant. See 37 CFR 1.313 and MPEP 1308.				
1. This communication is responsive to <u>10/13/05</u> .				
2.  The allowed claim(s) is/are <u>1-9,11-21 and 23-27</u> .				
3. Acknowledgment is made of a claim for foreign priority una a) All b) Some* c) None of the:  1. Certified copies of the priority documents have 2. Certified copies of the priority documents have 3. Copies of the certified copies of the priority documents have International Bureau (PCT Rule 17.2(a)).  * Certified copies not received:  Applicant has THREE MONTHS FROM THE "MAILING DATE" of noted below. Failure to timely comply will result in ABANDONM THIS THREE-MONTH PERIOD IS NOT EXTENDABLE.  4. A SUBSTITUTE OATH OR DECLARATION must be submit INFORMAL PATENT APPLICATION (PTO-152) which give	been received. been received in Application cuments have been received in this communication to file a ENT of this application.	No In this national stage application in this national stage application in the research of National Stage application in this national stage application in the research of the	quirements	
5. CORRECTED DRAWINGS (as "replacement sheets") mus  (a) including changes required by the Notice of Draftspers  1) hereto or 2) to Paper No./Mail Date  (b) including changes required by the attached Examiner's Paper No./Mail Date	on's Patent Drawing Review ( . Amendment / Comment or ir	n the Office action of		
Identifying indicia such as the application number (see 37 CFR 1.84(c)) should be written on the drawings in the front (not the back) of each sheet. Replacement sheet(s) should be labeled as such in the header according to 37 CFR 1.121(d).				
6. DEPOSIT OF and/or INFORMATION about the deposit attached Examiner's comment regarding REQUIREMENT F			Note the	
Attachment(s)  1. ☐ Notice of References Cited (PTO-892)  2. ☐ Notice of Draftperson's Patent Drawing Review (PTO-948)  3. ☑ Information Disclosure Statements (PTO-1449 or PTO/SB/0) Paper No./Mail Date 10/13/05  4. ☐ Examiner's Comment Regarding Requirement for Deposit	6. ☐ Interview Sun Paper No./M 8), 7. ⊠ Examiner's A	rmal Patent Application (PT nmary (PTO-413), ail Date mendment/Comment tatement of Reasons for Alle	ŕ	
of Biological Material				

9. Other \_\_\_\_.

## **EXAMINER'S AMENDMENT**

1. An examiner's amendment to the record appears below. Should the changes and/or additions be unacceptable to applicant, an amendment may be filed as provided by 37 CFR 1.312. To ensure consideration of such an amendment, it MUST be submitted no later than the payment of the issue fee.

Authorization for this examiner's amendment was given in a telephone interview with William Zychlewicz on 10/24/05.

2. The application has been amended as follows:

In claim 15, line 8, insert - - and - - after "phantom;".

## Reasons for Allowance

- 3. Claims 1-9, 11-21, and 23-27 are allowed. The following is an examiner's statement of reasons for allowance.
- 4. Regarding claim 1, prior art fails to disclose or fairly suggest a method including quantifying components of plaque by determining a weight sum of densities of pixels of an image of an organ including the plaque, in combination with all the limitations in the claim. Claims 2-9 and 11-13 are allowed by virtue of their dependency.
- 5. Regarding claim 14, prior art fails to disclose or fairly suggest a method including enabling visualization of restenosis within one of a metal stent and a metal valve by repeating

Art Unit: 2882

obtaining components of plaque after scanning a patient having the one of the metal stent and the

metal valve, in combination with all the limitations in the claim.

6. Regarding claim 15, prior art fails to disclose or fairly suggest a system including a

computer configured to repeat detection of components of plaque after instructing a user to

administer a temperature-sensitive contrast agent to accentuate a visualization of the plaque, in

combination with all the limitations in the claim. Claims 16-21 and 23-25 are allowed by virtue

of their dependency.

7. Regarding claim 26, prior art fails to disclose or fairly suggest a computer readable

medium encoded with a program configured to instruct a computer to repeat detection of

components of plaque after instructing a user to administer a contrast agent to accentuate a

visualization of the plaque, in combination with all the limitations in the claim.

8. Regarding claim 27, prior art fails to disclose or fairly suggest a computer encoded with a

program configured to quantify components of plaque by determining a weight sum of densities,

greater than a specific amount, of pixels of an image of an organ including the plaque, in

combination with all the limitations in the claim.

Any comments considered necessary by applicant must be submitted no later than the

payment of the issue fee and, to avoid processing delays, should preferably accompany the issue

Application/Control Number: 10/625,437

Art Unit: 2882

fee. Such submissions should be clearly labeled "Comments on Statement of Reasons for

Allowance."

Conclusion

Any inquiry concerning this communication or earlier communications from the

examiner should be directed to Chih-Cheng Glen Kao whose telephone number is (571) 272-

2492. The examiner can normally be reached on M - F (9 am to 5 pm).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's

supervisor, Ed Glick can be reached on (571) 272-2490. The fax phone number for the

organization where this application or proceeding is assigned is 571-273-8300.

Information regarding the status of an application may be obtained from the Patent

Application Information Retrieval (PAIR) system. Status information for published applications

may be obtained from either Private PAIR or Public PAIR. Status information for unpublished

applications is available through Private PAIR only. For more information about the PAIR

system, see http://pair-direct.uspto.gov. Should you have questions on access to the Private PAIR

system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free).

**∕** ok

EDWARD GLON

Page 4